



Evidence review:
Argus II retinal prosthesis for retinitis pigmentosa

NHS England

Evidence review: Argus II retinal prosthesis for retinitis pigmentosa

First published: December 2015

Updated: Not applicable

Prepared by: Ophthalmology CRG on behalf of NHS England Specialised
Commissioning

For public consultation

DRAFT FOR CLINICAL PANEL

Contents

1. Introduction	4
2. Summary of results.....	4
3. Research questions.....	4
4. Methodology	4
5. Results.....	5
6. Appendix One.....	8
7. Appendix Two.....	12

For public consultation

DRAFT FOR CLINICAL PANEL

1. Introduction

The Argus II retinal prosthesis system consists of video camera mounted on a pair of spectacles which communicates wirelessly with an implant placed surgically in the retina. The system stimulates the retina electrically with patterns which the brain is able to interpret as patterns of light.

2. Summary of results

The Argus II study group has studied a cohort of 30 patients and published several papers^{4,6-9,11}. A total of 29 of 30 subjects had functioning Argus II Systems implants 3 years after implantation. Eleven subjects experienced a total of 23 serious device- or surgery-related adverse events. All were treated with standard ophthalmic care. As a group, subjects performed better with the system on than off on all visual function tests and functional vision assessments¹¹ but there is variation between patients and between tests.

In the real world assessment (FLORA) three years after implantation, 65% of 23 patients were graded 'positive' or 'mildly positive' (i.e. usually subjects who self-reported functional benefits that were not supported by assessors' observations). An economic study¹⁰ modelled patients implanted at age 49 years with a 25 year time horizon. The incremental cost effectiveness ratio was 14 603 Euro (£10 634) per quality adjusted life year (QALY) for the base case, rising to 49 769 Euro (£36 206) with a shorter time horizon and different model assumptions. This study was populated with data from the Argus II cohort after two years of follow up.

3. Research questions

The research question is the effectiveness of the Argus II prosthesis in the treatment of visual loss due to retinitis pigmentosa.

Population: Patients with retinitis pigmentosa

Intervention: Argus II

Comparator: conventional standard of care

Outcome: Visual function, quality of life, adverse events.

4. Methodology

A search was undertaken on 25 August 2015 using PubMed, with 'Argus II retinal prosthesis' as the search term and 'clinical trial' as a filter:

(argus[All Fields] AND ii[All Fields] AND ("visual prosthesis"[MeSH Terms] OR

DRAFT FOR CLINICAL PANEL

("visual"[All Fields] AND "prosthesis"[All Fields]) OR "visual prosthesis"[All Fields] OR ("retinal"[All Fields] AND "prosthesis"[All Fields]) OR "retinal prosthesis"[All Fields]) AND Clinical Trial[ptyp]

One study of cost effectiveness was identified.

5. Results

The PubMed search retrieved nine papers¹⁻⁹. One study of cost effectiveness was identified by the PubMed suggestions system¹⁰. A further two papers were retrieved by citation tracking¹¹⁻¹².

The Argus II study group has studied a cohort of 30 patients over three years or more. Twenty-nine patients had retinitis pigmentosa and one subject had choroideremia. Twenty-nine patients had bare light perception (i.e., the ability to detect very bright light) in both eyes, and one subject had no light perception (but was able to perceive light in response to trans-corneal electrical stimulation). The age at time of implant ranged from 28 to 77 years (average 58 years).

Outcomes were measured in three ways:

- Computerised tests of square localization, direction of motion and grating visual acuity;
- Find the Door and Follow the Line; and
- Functional Low Vision Observer Rated Assessment (FLORA).

FLORA¹³ was developed specifically for this study at the request of the FDA to subjectively assess real-world benefit. Assessors interviewed the patient; observed the patient performing visual tasks (system on and off) in and around his or her home. The assessor then wrote a case study narrative to synthesize his or her judgment of the effect of the Argus II. All narratives were rated by a single independent rater for the effect of the system on subjects' lives: positive, mild positive (usually subjects who self-reported functional benefits that were not supported by assessors' observations), prior positive (subjects who self-reported positive effects in the past that could not be demonstrated at the time of the assessment), neutral, and negative.

On FLORA, 65% of 23 patients were rated positive or mildly positive (compared to previous) at three years.

On the functional tests, 89% of subjects performed significantly better with the system on than off for Square Localization, 56% for Direction of Motion, and 33% scored on the scale on Grating Visual Acuity with the system on (no subjects scored with the system off).

Seven subjects underwent elective revision surgeries in attempts to improve the position of the array.

As of 1 year after implantation, 66.7% of subjects (20/30) had experienced no

DRAFT FOR CLINICAL PANEL

device- or surgery-related SAEs. There were 18 SAEs among 10 subjects. The SAEs fell into 10 types, with hypotony, conjunctival dehiscence, conjunctival erosion, and presumed endophthalmitis (culture negative) being slightly more common than the others. There were also 2 subjects who underwent revision. At 3 years after implantation, there were a total of 23 SAEs among 11 subjects, with 2 additional SAE types. One subject's device was removed at 1.2 years to treat recurrent conjunctival erosion.

Other papers reported the results of various tests of visual function, including motion detection⁴, letter reading⁶, path tracing⁷, spatial tracking⁹, and navigation¹². Improvements varied from task to task, with 96% (26/27) of patients showing improved accuracy on the tracking task, 54% better on motion detection and equivocal results for the navigation task.

References

- 1: Rizzo S, Belting C, Cinelli L, Allegrini L. Visual field changes following implantation of the Argus II retinal prosthesis. *Graefes Arch Clin Exp Ophthalmol*. 2015 Feb;253(2):323-5. doi: 10.1007/s00417-014-2822-0. Epub 2014 Nov 30. PubMed PMID: 25432093.
- 2: Sabbah N, Authié CN, Sanda N, Mohand-Said S, Sahel JA, Safran AB. Importance of eye position on spatial localization in blind subjects wearing an Argus II retinal prosthesis. *Invest Ophthalmol Vis Sci*. 2014 Nov 20;55(12):8259-66. doi: 10.1167/iovs.14-15392. PubMed PMID: 25414187.
- 3: Luo YH, Davagnanam I, dacCuz L. MRI brain scans in two patients with the argus II retinal prosthesis. *Ophthalmology*. 2013 Aug;120(8):1711-1711.e8. doi:10.1016/j.ophtha.2013.04.021. PubMed PMID: 23916083.
- 4: Dorn JD, Ahuja AK, Caspi A, da Cruz L, Dagnelie G, Sahel JA, Greenberg RJ, McMahon MJ; Argus II Study Group. The Detection of Motion by Blind Subjects With the Epiretinal 60-Electrode (Argus II) Retinal Prosthesis. *JAMA Ophthalmol*. 2013 Feb;131(2):183-9. PubMed PMID: 23544203; PubMed Central PMCID: PMC3924899.
- 5: Ahuja AK, Behrend MR. The Argus™ II retinal prosthesis: factors affecting patient selection for implantation. *Prog Retin Eye Res*. 2013 Sep;36:1-23. doi: 10.1016/j.preteyeres.2013.01.002. Epub 2013 Mar 14. PubMed PMID: 23500412.
- 6: da Cruz L, Coley BF, Dorn J, Merlini F, Filley E, Christopher P, Chen FK, Wuyyuru V, Sahel J, Stanga P, Humayun M, Greenberg RJ, Dagnelie G; Argus II Study Group. The Argus II epiretinal prosthesis system allows letter and word reading and long-term function in patients with profound vision loss. *Br J Ophthalmol*. 2013 May;97(5):632-6. doi: 10.1136/bjophthalmol-2012-301525. Epub 2013 Feb 20. PubMed PMID: 23426738; PubMed Central PMCID: PMC3632967.
- 7: Barry MP, Dagnelie G; Argus II Study Group. Use of the Argus II retinal prosthesis to improve visual guidance of fine hand movements. *Invest Ophthalmol Vis Sci*. 2012 Aug 1;53(9):5095-101. doi: 10.1167/iovs.12-9536. PubMed PMID: 22661464; PubMed Central PMCID: PMC3416020.

DRAFT FOR CLINICAL PANEL

8: Humayun MS, Dorn JD, da Cruz L, Dagnelie G, Sahel JA, Stanga PE, Cideciyan AV, Duncan JL, Elliott D, Filley E, Ho AC, Santos A, Safran AB, Arditi A, Del Priore LV, Greenberg RJ; Argus II Study Group. Interim results from the international trial of Second Sight's visual prosthesis. *Ophthalmology*. 2012 Apr;119(4):779-88. doi: 10.1016/j.ophtha.2011.09.028. Epub 2012 Jan 11. PubMed PMID: 22244176; PubMed Central PMCID: PMC3319859.

9: Ahuja AK, Dorn JD, Caspi A, McMahon MJ, Dagnelie G, Dacruz L, Stanga P, Humayun MS, Greenberg RJ; Argus II Study Group. Blind subjects implanted with the Argus II retinal prosthesis are able to improve performance in a spatial-motor task. *Br J Ophthalmol*. 2011 Apr;95(4):539-43. doi: 10.1136/bjo.2010.179622. Epub 2010 Sep 29. PubMed PMID: 20881025; PubMed Central PMCID: PMC3345188.

10: Vaidya A et al. The cost-effectiveness of the Argus II retinal prosthesis in Retinitis Pigmentosa patients. *BMC Ophthalmol*. 2014 Apr 14;14:49. doi: 10.1186/1471-2415-14-49

11. Ho AC et al. Long term results from an epiretinal prosthesis to restore sight to the blind. *Ophthalmology* 2015; 122: 1547 – 54 doi: 10.1016/j.ophtha.2015.04.032. Epub 2015 Jul 8.

12. Garcia S, Petrini K, Rubin GS, da Cruz L, Nardini M. Visual and Non-Visual Navigation in Blind Patients with a Retinal Prosthesis. *PLoS One*. 2015; 10(7): e0134369.

13. Geruschat DR et al. FLORA: Phase I development of a functional visual assessment for prosthetic vision users. *Clin Exp Optom* 2015 DOI:10.1111/cxo.12242

DRAFT FOR CLINICAL PANEL

6. Appendix One

Grade of evidence	Study design			Outcomes				Reference	Other		
	Study design	Study size	Intervention	Primary Outcome	Primary Result	Secondary Outcome	Secondary Result	Reference	Complications noted	Benefits noted	Comments
2++	Case series	n=30	Argus II prosthesis					1			
2+	Cohort	n=13	Argus II prosthesis	Ability to detect motion	Fifteen subjects (54%) were able to perform the task significantly better with their prosthesis system than they were with their residual vision, 2 subjects had significantly better performance with their residual vision, and no difference was found for 11 subjects			4			
2+	Cohort	n=30	Argus II	Letter and word reading	21/30 patients in the cohort took part in the study. Results varied from letter to letter but for example for the letters L, T, E, J, F, H, I, U, patients recognized			6			

DRAFT FOR CLINICAL PANEL

					72% and 18% respectively with the system switched on or off.					
2+	Cohort		Argus II study group	Tracing paths on a touchscreen	For right-angle/single-turn sets, average tracing error was reduced by 63% and tracing time increased by 156% when using the prosthesis, relative to residual vision. With mixed-angle/single-turn sets, error was reduced by 53% and time to complete tracing increased by 184%. Prosthesis use decreased error by 38% and increased tracing time by 252% for paths that incorporated two turns			7		Not clear why only 21 patients were reported in this study from the Argus II group
	Cohort		Argus II study group Interim safety and efficacy	Safety (the number, seriousness, and relatedness of adverse	Subjects performed statistically better with system ON			8		

DRAFT FOR CLINICAL PANEL

			<p>results (see reference 11 for longer follow up)</p>	<p>events) and visual function, as measured by 3 computer-based, objective tests: square localization, direction of motion, grating visual acuity</p>	<p>vs. OFF in the following tasks: object localization (96% of subjects); motion discrimination (57%); and discrimination of oriented gratings (23%). The best recorded visual acuity to date is 20/1260. Subjects' mean performance on Orientation and Mobility tasks was significantly better when the System was ON vs. OFF. Seventy percent of the patients did not have any serious adverse events (SAEs). The most common SAE reported was either conjunctival erosion or dehiscence over the extraocular implant and was</p>						
--	--	--	--	---	---	--	--	--	--	--	--

DRAFT FOR CLINICAL PANEL

					successfully treated in all subjects except in one which required explanation of the device without further complications					
	Cohort	n=30	Argus II study cohort	Spatial tracking	Ninety-six percent (26/27) of subjects showed a significant improvement in accuracy and 93% (25/27) show a significant improvement in repeatability with the system on compared with off			9		
	Cohort	n=4 patients with Argus; n= 5 age matched and n= 6 younger controls		Path reproduction and a triangle completion navigation task	Results suggested no benefit on the path reproduction task; two patients showed benefit on the triangle task.			12		

DRAFT FOR CLINICAL PANEL

7. Appendix Two

Literature Search Terms

Assumptions / limits applied to search:	
Original search terms:	N/A
Updated search terms - Population	Patients with retinitis pigmentosa
Updated search terms - Intervention	Argus II
Updated search terms – Comparator	Supportive care.
Updated search terms – Outcome	<p><i>Critical to decision-making:</i></p> <p>Visual function, quality of life, adverse effects</p> <p><i>Important to decision-making:</i></p>
Inclusion criteria	General inclusion criteria
	Specific inclusion criteria
Exclusion criteria	General exclusion criteria
	Specific exclusion criteria